EU Quality Management System Certificate

We hereby certify the company

Falcon Medical Medizinische Spezialprodukte GmbH
Meiereigasse 2
2340 Mödling
Austria

the introduction and application of a quality management system in accordance with Annex IX, Chapter I and III of Regulation (EU) 2017/745 for conformity assessment.

An audit by mdc has proven that this quality management system meets the following requirements:

Annex IX – Chapter I (Quality Management System)

of Regulation (EU) 2017/745 of the European Parliament and of the Council of 5 April 2017 on medical devices.

Surveillance is carried out in accordance with Annex IX, Section 3 of Regulation (EU) 2017/745.

This certificate from mdc medical device certification GmbH (Notified Body 0483) consists of 2 pages. Details about the devices covered as well as further information and conditions are contained on the following pages.

Valid from	2024-08-26	Registration No.	D4007300010
Valid until	2028-08-22	Report No.	P24-00019-288110

Stuttgart, 2024-08-26

Notified Body

Benannt durch/Designated by
Zentralstelle der Länder für Gesundheitsschutz bei Arzneimitteln und Medizinprodukten
www.zlg.de
BS-MDR-098

Devices:

Reusable surgical instruments (trial components, broaches)

Risk class: IIa

Reusable surgical instruments

Risk class: I (reusable)

Artificial hip joint implant systems

Risk class: III

Notes:

In the case of class I devices that are reusable surgical instruments the involvement of mdc is limited to the aspects relating to the reuse of the device, in particular cleaning, disinfection, sterilization, maintenance and functional testing as well as the related instructions for use.

For the placing on the market of class III and IIb implantable devices (except sutures, staples, dental fillings, dental braces, tooth crowns, screws, wedges, plates, wires, pins, clips and connectors within the meaning of Regulation (EU) 2017/745, Art. 52 (4), 2nd paragraph and with the exception of custom-made devices of class III), an EU technical documentation assessment certificate is also required.

The certificate is based on the previous certificate

D4007300008 (2024-03-19)

with the following changes to D4007300008:
Supplemented by the product group "Artificial hip joint implant systems" Risk class III